

EFFECT OF BILIRUBIN CONCENTRATION ON RADIATION ABSORBED DOSE EQUIVALENT USING LiF (TLD-100) CHIPS

N. O. EGBE, B. F. OLISEMEKE, A. F. NNEOYI-EGBE, M. U. ETUNG, E. U. EYONG and R. R. ETTARH

(Received 6 October, 1999; Revision accepted 20 July, 2000)

ABSTRACT

The effect of bilirubin concentration on dose equivalent of absorbed radiation was investigated *in vitro* in this study. Different concentrations of bilirubin were exposed to a fixed dose equivalent x-ray irradiation from an R501 x-ray generator. The fraction of radiation transmitted and the fraction absorbed were measured using LiF (TLD 100) chips as detectors, and the radiation absorbed by bilirubin was computed. Result indicate that at low concentrations (25 $\mu\text{mol/L}$ to 76 $\mu\text{mol/L}$) absorbed doses decreased with increase in bilirubin concentration. At higher bilirubin concentrations (76 $\mu\text{mol/L}$ to 460 $\mu\text{mol/L}$) and beyond, there was an increase in absorption with a strong positive correlation ($r = 0.92$) between dose absorbed and bilirubin concentration. The possible applications of this property of bilirubin as a modifier at high concentrations- to enhance radiation effect on diseased tissue during radiotherapy, and the danger inherent presenting neonates for radiodiagnostic examinations are discussed.

Key words: Bilirubin, radiation, absorption, radiotherapy, modifier.

INTRODUCTION

Bilirubin is a final product of red blood cells breakdown. In the presence of NADPH and oxygen, the microsomal enzyme of the reticuloendothelial system, particularly the liver and spleen, adds a hydroxyl group to the methenyl bridge between two pyrrole rings with concomitant oxidation of ferrous ion- Fe^{2+} to ferric ion- Fe^{3+} (Voet and Voet, 1990). Further oxidation results in cleavage of the porphyrin ring with release of ferric iron and carbon monoxide to yield a green pigment biliverdin, which is reduced to bilirubin. The resulting bilirubin is transported in plasma bound to albumin (Champe and Harvey, 1994).

Due to the short span of the erythrocytes and increased haemoglobin mass, neonates have higher bilirubin levels than adults (Canter, 1989). Serum bilirubin levels are affected by both physiological and pathological conditions. They are elevated in both pregnancy and exercise (Eastham, 1985). Increased levels of total Bilirubin (conjugated and

unconjugated has diagnostic implications as it is associated with liver diseases and hemolytic disorders in adults and the newborn. Conjugated bilirubin is increased in biliary obstruction and hepatocellular disease (Eastham, 1985).

Serum bilirubin levels of patients are often requested before the radiographic examination of the renal system is performed. This is

Table 1 Doses of radiation transmitted and absorbed by different concentrations of bilirubin

Bilirubin concentration ($\mu\text{mol/L}$)	Dose transmitted (μSv)	Dose absorbed by bilirubin (μSv)
25.65	0.0044	0.0112
42.75	0.0066	0.0090
59.85	0.0105	0.0051
76.95	0.0124	0.0032
94.05	0.0113	0.0043
111.15	0.0083	0.0073
128.25	0.0066	0.0090
145.35	0.0050	0.0106
162.45	0.0022	0.0134
179.55	0.0017	0.0139

because excess bilirubin in the blood is known to dissolve contrast media (Mc Aurish, 1986). Apart from the contra-indication of high serum bilirubin level ($17\mu\text{mol/L}$ or 0.8mg/ml) for intravenous urography, there is paucity of data on the relationship between bilirubin concentration and absorbed radiation in tissues. This work investigates this relationship, and the possible application of this property in Radiotherapy.

MATERIALS AND METHODS

Lithium Fluoride thermoluminescent dosimeter chips (TLD-100) and pure bilirubin powder were obtained from the Sigma Chemical Company, England, and used for the experiments. An X-ray generator, model R501, with an output of 25 - 150 kV; 40 - 500 mA and 0.01 - 0.3 seconds, having a light beam diaphragm was used as the source of radiation, while the Lithium Fluoride thermoluminescent chips were used for recording transmitted doses as absorbed by the chips - therein read as absorbed doses (Gangadharan et al, 1986; Kitis et al, 1993).

Different concentrations of bilirubin ranging from $1.5\text{mg}/100\text{ml}$ to $10.5\text{mg}/100\text{ml}$, and corresponding to the approximate range of serum bilirubin levels in the different body compartments were prepared by dissolving weighed quantities of bilirubin powder in distilled water. The concentrations were expressed in $\mu\text{mol/L}$. The flasks were each labelled with the respective concentration and a control with distilled water only, set up. Each of the concentrations was exposed to x-ray irradiation of dose equivalent $0.01668\mu\text{Sv}$ and doses transmitted by the various concentrations and absorbed by the Lithium Fluoride chips were read off with a vinten Solaro dual channel TLD reader, model 680. The reader was operating at 160°C




Fig. 1 Plot of dose of radiation absorbed against bilirubin concentration

preheating temperature for 10 seconds, 300°C for 12 and 10 seconds read and anneal temperatures respectively. The dose absorbed by each of the different bilirubin concentrations was computed as follows for normal incidence with negligible reflection

$$\text{Dose absorbed} = \text{Incident radiation dose} - \text{Dose transmitted (TLD reading)}.$$

A plot of absorbed dose equivalent against bilirubin concentration was obtained to establish the relationship between both parameters.

Statistical Analysis

Results obtained were analyzed using Pearson's moment correlation coefficient to test the relationship between the two variables (Snedecor and Cochran, 1967).

RESULTS AND DISCUSSION

Table 1 presents results of the study showing transmitted dose, dose absorbed and the computed dose absorbed by the bilirubin. At high bilirubin concentrations ($>76\mu\text{mol/L}$), there was a strong positive correlation ($r = +0.92$) between bilirubin concentration and dose absorbed. At low concentrations of bilirubin ($25-75\mu\text{mol/L}$), there was a decrease in dose absorption, b

above $75\mu\text{mol/L}$, there was a continuous, though not uniform increase going up to $179.5\mu\text{mol/L}$, the highest bilirubin concentration used for this work.

The elemental composition of any medium traversed by radiation is an important factor governing the magnitude of interaction that occurs within the medium (McAurish, 1986). By implication every increase in the concentration of bilirubin (increase in elemental composition) should lead to an increase in the attenuation (absorption) of the X-ray beam. Our results show a departure from this theory at low concentrations of bilirubin. There was a marked and continuous decrease in radiation absorbed at concentrations between $25\mu\text{mol/L}$ and $5\mu\text{mol/L}$. Beyond $75\mu\text{mol/L}$ there was an increase in absorption. The sudden and unexpected "twist" in absorption suggests that bilirubin at concentrations below $76\mu\text{mol/L}$ may display an anomalous absorption characteristic.

Patients whose bilirubin concentrations fall within this range (25 - $76\mu\text{mol/L}$) may be adjudged "relatively safe" to undergo radiological examination, since a high percentage of the radiation dose in the primary beam is transmitted (Christense et al, 1978). This may not however imply a low degree of biological damage. Concentrations above $76\mu\text{mol/L}$ show a remarkable agreement with McAurish (1986). Increased concentrations led to a continuous increase in the absorbed doses as the concentration rose, suggesting that a greater percentage of the incident beam is absorbed. This may have a serious and grave consequence in the presence of oxygen, of which the body tissues are largely composed. However, since the direct effect of radiation is quantitatively more biologically harmful when compared to

the indirect action, (Yarminenko, 1988) it is suggested that bilirubin at high concentration may be a possible modifier in radiotherapy owing to its ability to absorb significant proportions of radiation.

When injected to cancerous tissue before radiotherapy, bilirubin may enhance radiation effects on the diseased cells, during treatment. It is suggested from the above that neonates (3 to 14 day old babies) and pregnant women should not be presented for radiodiagnostic examination unless there is no alternative to such an examination since at this age and during gestation, serum bilirubin levels are very high (1.0 to 10 mg/100 ml) according to Eastham (1985).

CONCLUSION

Bilirubin concentration displays significant radiation absorption *in vitro*, especially at high concentrations but its radiation absorption characteristics fail to agree with the results of McAurish (1986) as witnessed at lower concentrations.

REFERENCES

- Baker, F. J., 1985. Introduction to Medical Laboratory Technology. Butterworth and Co., London, 250p.
- Canter, D., 1989. Disease of the Gastrointestinal Tract and Liver. Churchill Livingstone, England. 617p.
- Champe, P. C. and Harvey, R. A., 1994. Lippincott's Illustrate Reviews: Biochemistry. 2nd ed. J. B. Lippincott Philadelphia: 262 - 264.
- Christensen, E., Curry, T.S. and Dardey J. E., 1978. An Introduction to Physics and

- Diagnostic Radiology. 2nd ed. Henry Kingston, London:450p.
- Eastham, R.D. 1985. Biochemical Values in Clinical Medicine. John Wright, England:789p
- McAurish, T. F. (ed) 1986. Physics in Medicine and Biology Encyclopedia. Vol. 2. Pergamon Press, Oxford:420 - 431, 618 -620.
- Sigma Chemicals, 1996. Diagnostic kits and Reagents. Sigma - Aldrich Co., Poole, U. K. 2363 -2366.
- Yarmonenko, S. P. 1988. Radiobiology of Humans and Animals. Mir Publishers, Moscow.:450p.
- Voet, D and Voet, J. G., 1990. Biochemistry. 1st ed. Singapore. John Wiley and sons. :890p.
- Gangadharan, P., Balasuramanian, V., Kale, L. R., Sane, S. G. and Pendurkar, H. K. 1968., TLD System for Measurement of wide range of X - and gamma Exposures. Proceedings of a Regional Seminar on Radiation Protection Monitoring. :307 - 319.
- Kitis, G., Charalambous, S. and Tuyn, J. W. N. 1993., Supralinearity of the thermoluminescent glow - peak V of LiF (TLD- 700)obtained with Isothermal measurements. J. Phys. D: Appl. Phys. 26 :2036 -2040
- Snedecor, G. W and Cockram, W. G 1967 Eds. Correlation In: Statistical Methods 6thed. Iowa State University Press, :172 - 198.