

Variant-specific RT-qPCR for rapid screening of B.1.617 mutations in SARS-CoV-2

Mwada Jallul, Khaled Ibrahim, Ahmed Zaghdani, Mohamed Musbah Abdusalam, Samira M Al Dwigen, Wafya S Atwair, Mohamed Elbasir, Inas Alhudiri, Salah Edin El Meshri & Adam Elzagheid

To cite this article: Mwada Jallul, Khaled Ibrahim, Ahmed Zaghdani, Mohamed Musbah Abdusalam, Samira M Al Dwigen, Wafya S Atwair, Mohamed Elbasir, Inas Alhudiri, Salah Edin El Meshri & Adam Elzagheid (2022) Variant-specific RT-qPCR for rapid screening of B.1.617 mutations in SARS-CoV-2, *Libyan Journal of Medicine*, 17:1, 2121252, DOI: [10.1080/19932820.2022.2121252](https://doi.org/10.1080/19932820.2022.2121252)

To link to this article: <https://doi.org/10.1080/19932820.2022.2121252>



© 2022 The Author(s). Published by Informa UK Limited, trading as Taylor & Francis Group.



Published online: 05 Sep 2022.



[Submit your article to this journal](#)



Article views: 239






[View related articles](#)



[View Crossmark data](#)

Variant-specific RT-qPCR for rapid screening of B.1.617 mutations in SARS-CoV-2

Mwada Jallul ^a, Khaled Ibrahim^a, Ahmed Zaghdani^b, Mohamed Musbah Abdusalam^a, Samira M Al Dwigen^c, Wafya S Atwair^a, Mohamed Elbasir^a, Inas Alhudiri ^a, Salah Edin El Meshri^b and Adam Elzagheid ^a

^aGenetic Engineering Department, Libyan Biotechnology Research Center, Tripoli, Libya; ^bDepartment of Microbiology, Libyan Biotechnology Research Centre, Tripoli, Libya; ^cDepartment of Cell Biology, Libyan Biotechnology Research Centre, Tripoli, Libya

ABSTRACT

The continuous emergence of new SARS-CoV-2 variants required rapid and reliable diagnostic methods for early detection and monitoring of the spread of the virus, especially in low-resource countries where whole genome sequencing is not available. We aimed to evaluate and compare the performance of two different RT-qPCR screening assays for the detection of B.1.617 lineage mutations. A total of 85 SARS-CoV-2 positive samples were collected between 9th August and 10 September 2021 and screened by two mutation-specific RT-qPCR assays for simultaneous detection of B.1.617.1 and B.1.617.2 lineage mutations. VIASURE Variant II PCR assay identified 2 Delta variant-specific mutations (L452R, and P681 R) in 80% of tested samples, while the PKamp™ Variant Detect™ assay was only able to detect one Delta variant specific mutation (L452R) in 75% of tested samples. This is the first report to show the Delta variant as the cause of the third wave in Libya. The use of multiplex RT-qPCR assays has allowed the identification of new variants for rapid screening. However, RT-qPCR results should be confirmed by whole genome sequencing of SARS-COV-2.

ARTICLE HISTORY

Received 27 April 2022
Revised 17 August 2022
Accepted 1 September 2022

KEYWORDS

RT- qPCR; B.1.617 lineage; delta variant; kappa variant; Libya

1. Introduction

Since the outbreak of COVID-19 in Wuhan, China, in 2019, several mutations have been shown in SARS CoV-2 [1,2]. From September 2020, new variants of concern (VOC) emerged in Europe [3], started from the UK, B.1.1.7 (alpha variant) [4], followed by two other variants: B.1.351 (beta variant) [5], and P.1 (gamma) [6]. In late 2020/early 2021 B.1.617 lineage was first identified in Maharashtra in India, and rapidly spreads around the globe [7]. The first sub-lineage to be detected was B.1.617.1 [7], followed by B.1.617.2, both carrying L452R mutation in the receptor binding domain of spike protein, which is also observed in Epsilon variant B.1.427/B.1.429 [7,8]. This mutation showed increased infectivity with little loss of susceptibility to neutralizing antibodies [9,10]. The mutations L452R, P681R, and E484Q within the receptor binding domain (RBD) are specific to the Kappa variant (lineage B.1.617.1) while L452R and P681R are specific to the Delta variant (lineage B.1.617.2), where P681R mutation is located in the S1-S2 furin cleavage site [11]. B.1.617.2, termed as Delta variant by WHO, has since dominated over B.1.617.1 (Kappa variant) and other lineages [12]. The symptoms of delta infection differ from pre-existing strains, usually starting with flu-like symptoms such as sore throat, headaches, and runny nose before progressing to pneumonia [13]. Because of the higher transmissibility and mortality

rate of the Delta variant, WHO alerted that this variant is rapidly becoming the main dominant strain of SARS-CoV-2 worldwide [14,15].

Since June 2021, the propagation of these variants in Libya required rapid screening based on RT-qPCR typing tests; because of their clinical implications of high transmissibility, and escaping from the immune system [16]. RT-qPCR is the gold standard for detection of SARS-CoV-2 in clinical samples [17], and has been used for the detection of many spike gene mutations focusing on deletion and SNPs in previous studies, using different techniques such as specific primers, and probes, and high-resolution melting curve analyses [18–24]

Here, we aimed to evaluate two mutation-specific RT-qPCR assays for screening of B.1.617 lineage mutations.

2. Material and methods

2.1. Samples collection

This study was conducted at the Molecular laboratory of the Libyan Biotechnology Research Center. A total of 85 clinical samples for SARS-CoV-2 positive patients were selected with a cycle threshold (Ct value) less than 35. All samples were previously tested N-gene positive by the Da An® detection kit for 2019-Novel Coronavirus (2019-nCoV) RNA (PCR-Fluorescence

Probing), V4, Da An® Gene Co., Ltd. of SunYat-sen University, which targeted the *ORF1ab* and *N* genes. The samples were collected from the Tripoli region in the period between 9th August and 10th September 2021.

Ethical approval was obtained from the Bioethics Committee of the Libyan Biotechnology Research Centre.

2.2. RNA extraction

RNA was extracted from the clinical samples using the Automated (Kingfisher flex Instrument, Thermo Fisher Scientific, Waltham, MA, USA), the kit used was viral Nucleic Acid (DNA/RNA) Extraction kit, Techstar, (WUXI TECH-STAR TECHNOLOGY CO., LTD) with 96 well pre-filled plates, following the manufacturer's instructions.

2.3. RT-qPCR and SARS-COV-2 variants detection

Positive samples were screened; on the same day, for the main mutations in lineage B.1.617, using two RT-qPCR variant-specific screening tests (VIASURE SARS-CoV-2 Variant II, Certest Biot, Zaragoza, Spain, and Pkamp™ VariantDetect™ SARS-CoV-2 RT-PCR assay, Combo F, PerkinElmer. Boston). The VIASURE SARS-CoV-2 Variant II assay screen for the presence of the *S* gene- E484Q, L452R, and P681R, whereas Pkamp™ VariantDetect™ screens for *S* gene- E484Q, L452R, E484K mutations, in addition to *N/ORF1ab*. Simultaneous detection of the L452R, E484Q, and P681R SNPs further allows differentiation between the B.1.617.1 and B.1.617.2 lineages in a single reaction.

The kits contained positive and negative control and Master-Mix reagents, including primers and probes for the detection of SARS CoV-2 targets. Briefly, VIASURE SARS-CoV-2 Variant II used 20 µl volume of reaction (15 µl M.M, and 5 µl sample), amplification was carried out with the following protocol: 45°C for 15 min, followed by 95°C for 2 min, and 45 cycles of 95°C for 10 sec, and 63°C for 50 sec. Another screening kit (Pkamp™ VariantDetect™ SARS-CoV-2 RT-PCR Kit, Combo F) used 30 µl volume of reaction (20 µl M.M, and 10 µl sample), with amplification protocol: 37°C for 2 min, followed by 50°C for 15 min, and 94°C for 10 min, and 45 cycles of 94°C for 10 sec, 62°C for 15 sec, and 65°C for 45 sec. following the manufacturer's recommendations. Amplification and detection were carried out on the Bio-Rad CFX96 Touch™ Real-Time PCR detection system, Bio-Rad (Bio-Rad, California, USA).

3. Results

The following RT-qPCR assays were primarily evaluated for their ability to detect the following mutations: E484K, E484Q, P6 81R, and L452R.

3.1. Pkamp™ Variant Detect™ SARS-CoV-2 RT-qPCR Kit

From the positive samples set, 75% of samples (n = 64) showed an L452R+/E484Q-/E484K - profile. Six samples were only positive for N and ORF 1ab gene, i.e. mutations in the *S* gene were not detected (L452R-/E484Q-/E484K-). The remaining 15 samples did not feature any of the targeted mutations.

3.2. VIASURE SARS-CoV-2 Variant II RT-PCR detection kit

Unlike Pkamp™ VariantDetect™, VIASURE SARS-CoV-2 Variant II assay additionally detects the presence of P681R mutation characteristic of B.1.617.1 and B.1.617.2 lineages. Sixty-eight (80%) samples had a P681R+/L452R+/E484Q- profile, which is consistent with the Delta variant. Only three cases (3.5%) showed a P681R+/L452R+/E484Q+ profile which is consistent with Kappa. No mutation specific to the Epsilon variant was detected. Furthermore, 14 samples showed no amplification curves for any targeted mutations.

3.3. Comparison of the two assays

Of note, three samples that were consistent with kappa variant (P681R+/L452R+/E484Q+) by VIASURE SARS-CoV-2 Variant II were only positive for L452R by Pkamp™ VariantDetect™ RT-qPCR detection kit. Pkamp™ VariantDetect™ kit does not target P681R mutation; hence the variant type cannot be determined based on E484K and E484Q alone, this decreases the kit specificity in comparison to VIASURE SARS-CoV-2 Variant II. A Comparison of results obtained by the two assays is summarized in Table 1. The difference between the two means of Ct value of L452R was statistically significant (P < 0.0001).

Collectively, Lineages B.1.617.2 (Delta variant) appeared to be predominant in positive samples. Interestingly, lineages B.1.427/B.1.429 (Epsilon variant) was not detected. VIASURE SARS CoV-2 Variant II was able to detect E484Q in three samples but Pkamp™ VariantDetect™ didn't. CT values of all mutations

Table 1. Comparison of results obtained by Pkamp™ VariantDetect™ (Combo F) and VIASURE SARS-CoV-2 Variant II assays for the detection of B.1.617 mutation analyzed samples (n = 85).

Type of Mutation	PKamp™ VariantDetect™ (Combo F)		VIASURE SARS CoV-2 Variant II	
	# Of positive cases	C _T (Mean)	# Of positive cases	C _T (Mean)
L452R	64	33.23	68	25.7
E484Q	ND		3	21
E484K	ND		NA	
P681R	NA		68	26.5

ND – Not Detected, NA- Not Applicable.




Figure 1. Dot plot of cycle thresholds in the two assays.

obtained by both assays are represented as dot plots (Figure 1).

4. Discussion

Rapid screening of SARS CoV-2 mutations can early identify circulating variants, may alert for early infections, rapid infection control, and public health interventions [25]. This study has evaluated two multiplex RT-qPCR-assays for rapid screening of newly circulating SARS-CoV-2 Variants. We identified 75% of cases with an L452R+/E484Q-/E484K- profile using Pkamp™ VariantDetect™ assay which detects three targets (N and ORF1ab genes, and E484K, E484Q, L452R in S-gene). Importantly, the variant couldn't be specified as Delta, since L452R mutation is also present in other variants (e.g. B.1.427&B.1.429) [26]. It is noteworthy, that six samples were only positive for N and ORF1ab but not for S gene mutations, four of these samples showed Ct above 35 for L452R mutation with VIASURE SARS-CoV-2Variant II assay. This may be due to the relatively higher Limit of detection of Pkamp™ VariantDetect™ (200 copies per PCR reaction) in comparison to 160 genome copies/reaction of VIASURE SARS CoV-2Variant II assay.

All samples used in this study were selected with Ct value ≤ 35 for the N gene using Da An® detection kit for 2019-Novel Coronavirus (2019-nCoV) RNA (PCR-Fluorescence Probing), V4, Da An® Gene Co., Ltd. of SunYat-sen University and the majority had Ct values <30 . Although N gene has greater sensitivity for detection of SARS-CoV-2 since it is abundant among other gene targets [27,28], some samples showed negative results for N gene using Pkamp™ VariantDetect™; this may suggest lower sensitivity of this kit. Furthermore, multiplex PCR assays using many target genes although have many advantages, they can cause self-inhibition among different sets of primers; low amplification efficiency; and variable efficiency in different templates [29].

After the screening for S gene mutations many samples showed a relatively low viral load (Ct values

≥ 31) using both kits, this may be due to low sensitivity of both assays. Since the Ct values for all samples were taken to be less than 35 then the issue of lower DNA concentration could be excluded.

Samples that showed no amplification with any target genes (15 samples with Pkamp™ VariantDetect™, 14 samples with VIASURE SARS CoV-2 Variant II) need sequencing since they might belong to other variants. Pkamp™ VariantDetect™ has lower efficiency to amplify S gene, since the Ct value of L452R mutation was higher than VIASURE SARS-CoV-2Variant II. Notably, these samples tested positive for the N-gene and ORF1ab with Da An gene Kit which is routinely used in our laboratories for SARS-COV-2 diagnosis and has a high sensitivity for N gene [30] and then screened for using Pkamp™ VariantDetect™, and VIASURE SARS CoV-2 Variant II on the same day of detection.

The detection of L452R mutation in the samples without any combination with other S gene mutations is meaningless for variant specification; because L452R is found in many circulating variants and its use for screening cannot determine which variant has infected the patient [26]; therefore it is crucial to differentiate between all samples that contain L452R, either through multiplex PCR assays targeting specific mutations commonly found along with L452R or by sequencing of S gene.

It is noteworthy, that T478K mutation or its combination with other mutations in spike gene showed a greater fitness to Delta variant [11], therefore its detection could discriminate between B.1.617 sub-lineages. Both RT-PCR kits used in this study lack T478K detection and since delta variant was the most dominant variant at the time of the study the detection of T478K in combination with other mutations is very important to avoid misdiagnosis.

As in many other countries, the delta variant wave began to spread around to many countries [13], Libya was no exception, in late June a sharp increase in the number of positive samples was attributed to Delta, this was in agreement with our results because the frequency of Delta cases was dominant (using VIASURE SARS CoV-2Variant II assay).

The unique combination of spike mutations L452R, E484Q, P681R has been observed in B.1.617.1 and B.1.617.3 in Maharashtra, India [11], this was consistent with our findings. Cherian and colleagues showed that the combination of two RBD mutations, L452R and E484Q, could affect the neutralization by specific monoclonal Antibodies [11]; therefore, the rapid detection of L452R and E484K/Q mutations is necessary; because this confers clinical implications on the vaccine and COVID-19 treatment; since these mutations associated with reduced efficacy of the vaccine, or confers resistance to neutralizing antibodies [31].

Several publications have evaluated the potential performance of different variant-specific PCR screening kits in comparison to whole genome sequencing using next-generation sequencing (NGS) technology. The authors confirmed a high concordance between the two methods, and demonstrate their varying degrees of suitability for practical use in diagnostics and screening methods [19,24,32–35]. However, the misidentification of variants by RT-qPCR which was then confirmed to belong to other variants by sequencing was reported by Boudet and colleagues [36].

The difficulty for accessing sequencing reagents and their costs are the main challenges for performing whole genome sequencing in Libya, in addition to the need for capacity building on bioinformatics and genomic data analysis.

Furthermore, continuous development of these PCR assays is important since SARS CoV-2 is continuously mutating resulting in the emergence of new variants for example Omicron which could be quickly detected by identifying its new spike mutations (i.e. ins214EPE, S371L, and S373P) [37].

Multiplex RT-qPCR assays demonstrate practical results to specifically screen for B.1.617 lineage variants (delta/kappa) [38], this was consistent with our results since VIASURE SARS-CoV-2 Variant II assay was able to detect E484Q mutation, and hence variant classification (according to manufacturer instructions).

5. Conclusion

To summarize, the results obtained from RT-qPCR screening kits represent an ideal balance of costs, turnaround time, and preliminary call for variants, and it is simple for implementation in the routine clinical laboratory. Ultimately, the definitive answer should only be made based on whole genome sequencing.

Acknowledgments

The authors acknowledge all staff members who worked at the laboratory for detection of SARS CoV-2 at Biotechnology Research Centre, Tripoli, Libya.

Disclosure statement

No potential conflict of interest was reported by the author(s).

Funding

Libyan Biotechnology Research Centre, Tripoli, Libya.

ORCID

Mwada Jallul  <http://orcid.org/0000-0001-6897-5785>
Inas Alhudiri  <http://orcid.org/0000-0002-3371-0215>
Adam Elzagheid  <http://orcid.org/0000-0002-7086-5937>

References

- [1] Baric RS, Phimister EG. Emergence of a highly fit SARS-CoV-2 variant. *N Engl J Med.* 2020;383:2684–2686.
- [2] Chen N, Zhou M, Dong X, et al. Epidemiological and clinical characteristics of 99 cases of 2019 Novel Coronavirus Pneumonia in Wuhan, China: a descriptive study. *Lancet.* 2020;395:507–513.
- [3] Virological. Preliminary genomic characterisation of an emergent SARS-CoV-2 lineage in the UK defined by a novel set of spike mutations - SARS-CoV-2 Coronavirus/ NCoV-2019 genomic epidemiology. Available from: <https://virological.org/t/preliminary-genomic-characterisation-of-an-emergent-sars-cov-2-lineage-in-the-uk-defined-by-a-novel-set-of-spike-mutations/563>. Accessed 2021 February 28.
- [4] Davies NG, Abbott S, Barnard RC, et al. Estimated transmissibility and impact of SARS-CoV-2 lineage B.1.1.7 in England. *Science.* 2021;372:eabg3055.
- [5] Tegally H, Wilkinson E, Giovanetti M, et al. Emergence and rapid spread of a new severe acute respiratory syndrome-related Coronavirus 2 (SARS-CoV-2) lineage with multiple spike mutations in South Africa. *medRxiv.* 2020. DOI:10.1101/2020.12.21.20248640.
- [6] Faria NR, Mellan TA, Whittaker C, et al. Genomics and epidemiology of the P.1 SARS-CoV-2 lineage in Manaus, Brazil. *Science.* 2021;372:815–821.
- [7] Ferreira IATM, Kemp SA, Datir R, et al. SARS-CoV-2 B.1.617 mutations L452R and E484Q are not synergistic for antibody evasion. *J Infect Dis.* 2021;224:989–994.
- [8] Deng X, Garcia-Knight MA, Khalid MM, et al. Transmission, infectivity, and antibody neutralization of an emerging SARS-CoV-2 variant in California carrying a L452R spike protein mutation. *medRxiv.* 2021. DOI:10.1101/2021.03.07.21252647.
- [9] McCallum M, Bassi J, De Marco A, et al. SARS-CoV-2 immune evasion by variant B.1.427/B.1.429. *bioRxiv.* 2021. DOI:10.1101/2021.03.31.437925.
- [10] Motozono C, Toyoda M, Zahradnik J, et al. An emerging SARS-CoV-2 mutant evading cellular immunity and increasing viral infectivity. *bioRxiv.* 2021. DOI:10.1101/2021.04.02.438288.
- [11] Cherian S, Potdar V, Jadhav S, et al. SARS-CoV-2 spike mutations, L452R, T478K, E484Q and P681R, in the second wave of COVID-19 in Maharashtra, India. *Microorganisms.* 2021;9:1542.
- [12] Hadfield J, Megill C, Bell SM, et al. Nextstrain: real-time tracking of pathogen evolution. *Bioinformatics.* 2018;34:4121–4123.
- [13] Essa RA, Ahmed SK, Bapir DH, et al. Clinical features and laboratory findings first case of B. 1.617.2 (delta) variant concern (VOC) in Iraq. *Ann Med Surg.* 2021;69:102814.
- [14] Lovelace B. WHO says delta is becoming the dominant covid variant globally, CNBC. WHO; 2021.
- [15] Sah R, Khatiwada AP, Shrestha S, et al. COVID-19 vaccination campaign in Nepal, emerging UK variant and futuristic vaccination strategies to combat the ongoing pandemic. *Travel Med Infect Dis.* 2021;41:102037.
- [16] Bal A, Destras G, Gaymard A, et al. Two-step strategy for the identification of SARS-CoV-2 variant of concern 202012/01 and other variants with spike deletion H69-V70, France, August to December 2020. *Euro Surveill.* 2021;26. DOI:10.2807/1560-7917.ES.2021.26.3.2100008.
- [17] Kevadiya BD, Machhi J, Herskovitz J, et al. Diagnostics for SARS-CoV-2 infections. *Nat Mater.* 2021;20:593–605.

- [18] Kováčová V, Boršová K, Paul ED, et al. Surveillance of SARS-CoV-2 lineage B.1.1.7 in Slovakia using a novel, multiplexed RT-QPCR assay. medRxiv. 2021. DOI:10.1101/2021.02.09.21251168.
- [19] Vega-Magaña N, Sánchez-Sánchez R, Hernández-Bello J, et al. RT-QPCR assays for rapid detection of the N501Y, 69-70del, K417N, and E484K SARS-CoV-2 mutations: a screening strategy to identify variants with clinical impact. Front Cell Infect Microbiol. 2021;11:672562.
- [20] Ratcliff J, Nguyen D, Fish M, et al. Virological and serological characterization of critically ill patients with COVID-19 in the UK: a special focus on variant detection. medRxiv. 2021. DOI:10.1101/2021.02.24.21251989.
- [21] Lemmermann NA, Lieb B, Laufs T, et al. SARS-CoV-2 genome surveillance in Mainz, Germany, reveals convergent origin of the N501Y spike mutation in a hospital setting. medRxiv. 2021. DOI:10.1101/2021.02.11.21251324.
- [22] Bier C, Edelmann A, Theil K, et al. Multi-site evaluation of SARS-CoV-2 spike mutation detection using a multiplex real-time RT-PCR assay. medRxiv. 2021. DOI:10.1101/2021.05.05.21254713.
- [23] Sandoval Torrientes M, Castelló Abietar C, Boga Riveiro J, et al. A novel single nucleotide polymorphism assay for the detection of N501Y SARS-CoV-2 variants. J Virol Methods. 2021;294:114143.
- [24] Vogels CBF, Breban MI, Ott IM, et al. Multiplex QPCR discriminates variants of concern to enhance global surveillance of SARS-CoV-2. PLOS Biol. 2021;19:e3001236.
- [25] Hale R, Crowley P, Dervisevic S, et al. Development of a multiplex tandem PCR (MT-PCR) assay for the detection of emerging SARS-CoV-2 variants. Viruses. 2021;13:2028.
- [26] Aoki A, Adachi H, Mori Y, et al. A rapid screening assay for L452R and T478K spike mutations in SARS-CoV-2 delta variant using high-resolution melting analysis. J Toxicol Sci. 2021;46:471–476.
- [27] Ogando NS, Dalebout TJ, Zevenhoven-Dobbe JC, et al. SARS-Coronavirus-2 replication in vero E6 cells: replication kinetics, rapid adaptation and cytopathology. J Gen Virol. 2020;101:925–940.
- [28] Ong DSY, Claas ECJ, Breijer S, et al. Comparison of the GeneFinder™ COVID-19 plus RealAmp kit on the sample-to-result platform ELITE InGenius to the national reference method: an added value of N gene target detection? J Clin Virol. 2020;132:104632.
- [29] Xu W, Shang Y. The detection techniques of genetically modified organisms. Genet Modif Org Food. 2016;343–351. DOI:10.1016/B978-0-12-802259-7.00031-2.
- [30] ElJilani M, Abdusalam M, Abdalla A, et al. Performance evaluation of seven commercial real-time polymerase chain reaction assays for SARS-COV-2 detection. Libyan J Med Sci. 2021;5:56.
- [31] Ferreira I, Datir R, Papa G, et al. SARS-CoV-2 B.1.617 emergence and sensitivity to vaccine-elicited antibodies. bioRxiv. 2021. DOI:10.1101/2021.05.08.443253.
- [32] Kami W, Kinjo T, Arakaki W, et al. Rapid and simultaneous identification of three mutations by the Novaplex™ SARS-CoV-2 variants i assay kit. J Clin Virol. 2021;141:104877.
- [33] Wang H, Miller JA, Verghese M, et al. Multiplex SARS-CoV-2 genotyping reverse transcriptase PCR for population-level variant screening and epidemiologic surveillance. J Clin Microbiol. 2021;59. DOI:10.1128/JCM.00859-21.
- [34] Amato L, Jurisic L, Puglia I, et al. Multiple detection and spread of novel strains of the SARS-CoV-2 B.1.177 (B.1.177.75) lineage that test negative by a commercially available nucleocapsid gene real-time RT-PCR. Emerg Microbes Infect. 2021;10:1148–1155.
- [35] Wang H, Jean S, Eltringham R, et al. Mutation-specific SARS-CoV-2 PCR screen: rapid and accurate detection of variants of concern and the identification of a newly emerging variant with spike L452R mutation. J Clin Microbiol. 2021;59. DOI:10.1128/JCM.00926-21.
- [36] Boudet A, Stephan R, Bravo S, et al. Limitation of screening of different variants of SARS-CoV-2 by RT-PCR. Diagnostics. 2021;11:1241.
- [37] De Pace V, Bruzzone B, Orsi A, et al. Comparative analysis of five multiplex RT-PCR assays in the screening of SARS-CoV-2 variants. Microorganisms. 2022;10:306.
- [38] Nörz D, Grunwald M, Tang HT, et al. Rapid automated screening for SARS-CoV-2 B.1.617 lineage variants (Delta/Kappa) through a versatile toolset of QPCR-based SNP detection. Diagnostics. 2021;11:1818.