




# A pandemic guided by the SARS-CoV-2 PCR test: What should the clinician know?

**Authors:**

Avania Bangalee<sup>1,2</sup>   
Kreshalen Govender<sup>1,2</sup>   
Varsha Bangalee<sup>3</sup> 

**Affiliations:**

<sup>1</sup>Department of Medical Virology, Faculty of Health Sciences, Prins Hof Campus, University of Pretoria, South Africa

<sup>2</sup>National Health Laboratory Services, Johannesburg, South Africa

<sup>3</sup>Discipline of Pharmaceutical Sciences, Faculty of Health Sciences, University of KwaZulu-Natal, Durban, South Africa

**Corresponding author:**

Varsha Bangalee,  
bangalee@ukzn.ac.za

**Dates:**

Received: 18 Jan. 2022

Accepted: 14 May 2022

Published: 15 Sept. 2022

**How to cite this article:**

Bangalee A, Govender K, Bangalee V. A pandemic guided by the SARS-CoV-2 PCR test: What should the clinician know? *S Afr Fam Pract.* 2022;64(1), a5492. <https://doi.org/10.4102/safp.v64i1.5492>

**Copyright:**

© 2022. The Authors. Licensee: AOSIS. This work is licensed under the Creative Commons Attribution License.

**Read online:**

Scan this QR code with your smart phone or mobile device to read online.

Amidst an ever-evolving pandemic, the demand for timely and accurate diagnosis of coronavirus disease 2019 (COVID-19) continues to increase. Critically, managing and containing the spread of the disease requires expedient testing of infected individuals. Presently, the gold standard for the diagnosis of severe acute respiratory syndrome coronavirus 2 (SARS-CoV-2) infection remains the polymerase chain reaction (PCR) test. Potential vulnerabilities of this testing methodology can range from preanalytical variables to laboratory-related analytical factors and, ultimately, to the interpretation of results.

**Keywords:** SARS-CoV-2 PCR, pandemic, COVID-19 testing, molecular diagnostic test, laboratory, COVID-19 response.

## Introduction

The coronavirus disease 2019 (COVID-19) pandemic represents a rapidly evolving public health emergency. The emergence of viral variants, a lack of diagnostic equipment and expertise coupled with inequitable vaccine access have resulted in global outbreaks. Consequently, the demand for testing continues to increase. The gold standard for the diagnosis of severe acute respiratory syndrome coronavirus 2 (SARS-CoV-2) infection remains the polymerase chain reaction (PCR) test. This test, however, requires skilled staff, laboratory infrastructure and specialised equipment. Moreover, there are numerous factors that influence the outcome and interpretation of a SARS-CoV-2 PCR test. This article will highlight potential pitfalls to consider when using this testing methodology.

## Pre-analytical factors to consider

The molecular diagnosis of COVID-19 rests on the detection of viral ribonucleic acid (RNA) in a clinical specimen. Infection begins in the upper respiratory tract (URT) and may progress to the lower respiratory tract (LRT) in more severe cases.<sup>1</sup> The most common method used for respiratory specimen collection is the nasopharyngeal swab (NPS). Another specimen that may be collected is the oropharyngeal swab; however, data suggests that it is slightly less sensitive than the NPS.<sup>2,3,4</sup> Saliva has been evaluated in various studies globally as it is a simple and non-invasive sample type that may be taken by a patient at home, thereby limiting contact with healthcare workers and the need for costly personal protective equipment. Moreover, saliva does not require expensive collection supplies and appears stable under variable conditions for a prolonged period.<sup>5</sup> A recent systematic review of the diagnostic performance of different sampling approaches for SARS-CoV-2 reverse transcriptase polymerase chain reaction (RT-PCR) testing found that combined nasal and throat swabs gave the highest sensitivity of 97%. Lower sensitivities were obtained from saliva specimens (85%) and nasal swabs (86%) with the lowest sensitivity achieved from throat swabs (68%).<sup>6</sup>

The Centre for Disease Control (CDC) and the South African National Department of Health have also recently approved mid-turbinate and nasal swabs for testing.<sup>7,8</sup> However, both these sample types are not routinely used in South Africa (SA) for PCR testing.

Suboptimal sample collection influences the amount of RNA detected. Specimen collection devices should have aluminium or plastic shafts as wooden shafts may contain substances that inactivate some viruses and inhibit PCR testing. The World Health Organization (WHO) recommends synthetic-tipped swabs such as rayon or nylon for SARS-CoV-2 nucleic acid detection<sup>9</sup>; however, shortage of collection supplies has led to the use of cotton swabs, especially in poorly resourced healthcare sectors. Whilst a recent study found that cotton does not inhibit

the detection of SARS-CoV-2 RNA, further studies are required to confidently confirm this.<sup>10</sup>

Operator technique is another critical factor that has often contributed to false-negative results. Improper technique results in swabs not reaching the target site; the aim is to brush infected cells and secretions off the nasopharyngeal mucous membranes. This ensures that sufficient RNA is obtained from the site of initial viral replication. There has been an influx in the number of drive-through sites at which tests are being performed to improve convenience and increase the number of tests done whilst adhering to social distancing recommendations. Despite the advantages offered by a drive-through model, literature interrogating the accuracy on SARS-CoV-2 PCR results is scant, as patient and operator positioning may not be optimised for proper NPS technique.<sup>11</sup>

Sample type may also affect a PCR result. Patients with evidence of LRT disease may require an LRT sample, especially if the patient is tested later in the course of illness.<sup>12</sup> Lower respiratory tract specimens include expectorated sputum, bronchoalveolar lavage (BAL) or endotracheal aspirates and should be submitted in clean universal containers without transport medium. These samples are especially useful when the pretest probability of disease is high and a URT specimen yielded a negative PCR result. A study evaluating the detection of SARS-CoV-2 RNA in a range of sample types found that BAL specimens showed the highest positivity rates (93%) followed by sputum (72%).<sup>13</sup>

Transport and storage of samples may also affect the integrity of viral RNA.<sup>14</sup> Ideally, samples should reach the laboratory as soon as possible and arrive in viral or universal transport medium at 2 °C – 8 °C. Dry swabs in a sterile tube may also be submitted. Dry swabs can be sent at ambient temperature but should reach the laboratory within 48 h. Unfortunately, testing delays and backlogs have led to widespread fluctuations in transport and storage conditions. A United States-based study examining the effect of extended storage at ambient temperature on NPS found minimal and clinically insignificant impact on PCR results.<sup>15</sup> Research is ongoing for molecular testing in lower-income settings under various storage conditions. In the interim, it is crucial to abide by the recommended storage and transport conditions to maintain sample integrity.

Timing of testing in relation to disease course is critical when submitting a sample to the laboratory. Viral load in the nasopharynx is highest in the immediate presymptomatic and symptomatic phase of the illness and decreases from week 3 to eventually become undetectable.<sup>15</sup> Therefore, an NPS taken too early or too late in relation to clinical symptoms may cause a false negative result. Other pre-analytical errors such as sample swaps, mislabelling and contamination may also contribute to false PCR results.

## In the laboratory: Analytical factors affecting a polymerase-chain reaction test

A SARS-CoV-2 PCR test is a very sensitive molecular method of reverse transcribing SARS-CoV-2 RNA into DNA, followed by amplification of the target genomic sequence. Fluorogenic probes are used to search for the target DNA sequence which generates a fluorescent signal that increases proportionally to the amount of viral RNA present in the sample. A cycle threshold (Ct) value is obtained when the fluorescence reaches a specific threshold within a certain number of PCR cycles. A Ct value less than 40 for one or more viral gene segments is generally reported as PCR positive. There is generally an inverse correlation between Ct values and viral loads, such that lower Ct values represent higher viral RNA loads.<sup>16</sup>

Specific SARS-CoV-2 gene segments are targeted by different assays. Generally, most PCR assays target two or three of the envelope (*env*), nucleocapsid (*N*), spike (*S*), RNA-dependent RNA polymerase (*RdRp*) or *ORF1* genes.<sup>17</sup> For any sample, these targets should take roughly the same number of PCR cycles to be detected. The Foundation for Innovative New Diagnostics (FIND) is compiling a database of SARS-CoV-2 molecular assays that are commercially available or in development for the diagnosis of COVID-19, accessible at [https://www.finddx.org/covid-19/pipeline/?section=molecular-assays#diag\\_tab](https://www.finddx.org/covid-19/pipeline/?section=molecular-assays#diag_tab). Currently, within the National Health Laboratory Service in SA, the six most widely used testing kits are presented in Table 1.

The PCR results are usually available within 24 h to 48 h of receipt and depend on patient priority (ill, hospitalised patients and healthcare workers are prioritised), the distance between the sample collection site and the testing laboratory, the volume of tests received at the laboratory and the assay(s) used.

Contamination has been a major thorn in the side of molecular assays and may be derived from two sources: cross-contamination between specimens or synthetically derived target nucleic acid.<sup>18</sup> Cross-contamination from a positive sample to a negative one may occur during sample preparation for PCR. This risk increases substantially during peaks of infection when viral loads in samples are very high. During the PCR test, billions of copies of nucleic acid are generated,

**TABLE 1:** SARS-CoV-2 molecular tests in South African public sector laboratories.

SARS-CoV-2 test	Company	Genes targeted
Alinity m	Abbott Laboratories, Abbott Park, IL, US	N, RdRp
Abbott RealTime	Abbott Laboratories, Abbott Park, IL, US	N, RdRp
Xpert <sup>®</sup> Xpress	Cepheid, Sunnyvale, CA, US	E, N
cobas <sup>®</sup>	Roche Molecular Systems Inc., Branchburg, NJ, US	ORF1, E
TaqPath™	ThermoFisher Scientific, Waltham, MA, US	S, E, N
Allplex™	Seegene Inc., Seoul, South Korea	E, N, RdRp

N, nucleocapsid; RdRp, RNA-dependent RNA polymerase; E, envelope; ORF1, open reading frame 1; S, spike; SARS-CoV-2, severe acute respiratory syndrome coronavirus 2; US, United States.

and these can contaminate instruments, reagents and samples. A lesser source of contamination is assay-derived, which can cause false positives and reduce the specificity of the diagnostic assay.<sup>19</sup> Whilst quality assurance steps such as the use of extraction controls and negative template controls may help to detect contamination, low-level contamination may go unnoticed. This scenario can occur during a peak of infection when there are many samples containing millions of copies of viral RNA, coupled with an urgency in the laboratory to ensure that samples are processed rapidly.

The SARS-CoV-2 has a high rate of error-prone replication and penchant for recombination.<sup>20</sup> As the pandemic has progressed, several genetically unique variants of concern, monitored by the WHO, have emerged globally, namely, B.1.1.7 (Alpha), B.1.351 (Beta), B.1.617.2 (Delta) and P.1 (Gamma).<sup>21</sup> Because molecular tests detect a specific viral nucleic acid sequence, a mutation may affect the performance of a PCR test if it occurs in a region of the genome targeted by the test. The impact of these variants on test performance is influenced by the design of the test, the sequence of the variant and the prevalence of the variant in the population. For example, the alpha variant carries a double deletion at positions 69 and 70 on the spike protein gene (S-gene). This appears to impact detection using an assay that targets the S-gene, such as the TaqPath™ COVID-19 Combo Kit.<sup>22</sup> This S-gene target failure was also observed with the recent Omicron variant which contained the same mutation. Fortunately, even if a mutation impacts one of the PCR targets, most commercial PCR-based tests have two or more targets to detect SARS-CoV-2. In the case of a mutation, one may consider retesting using a test targeting different genes if COVID-19 is still suspected after receiving a negative PCR test result.

## The postanalytical phase: Interpretation of a polymerase chain reaction result

If the PCR assay detects all viral gene targets (usually two or three), then viral RNA is detected and the test is positive; if no targets are detected, then viral RNA is absent from the sample and it is reported as a negative result; if only some of the targets are present, then the laboratory takes a decision based on an agreed Ct cut-off value whether to report this result as positive or as inconclusive. An inconclusive result can be the most problematic to troubleshoot. Polymerase chain reaction-based assays can yield a weak signal or nonspecific result near the limit of detection of the assay. This high Ct value generally indicates very little viral RNA in the sample.<sup>23</sup> This may be a result of a sample taken very early in the infection, intermittent viral shedding – usually at the tail-end of infection – or contamination.

It is important to remember that PCR is an extremely sensitive method and under optimal conditions can detect fewer than 10 copies of viral RNA in a clinical sample; however, it cannot distinguish between viable virus and noninfectious RNA. Virus viability appears to be brief with viral culture negative in cases with a Ct higher than 33.<sup>24,25,26</sup> If one correlates Ct values

with clinical progression, Ct values are lowest in the immediate presymptomatic and first week of the symptomatic phases of COVID-19.<sup>27</sup> Therefore, Ct values correlate closely with a clinical history and examination. Unfortunately, these details are often omitted on a request form, so that interpretation of a result is less case-specific. Moreover, PCR can often detect viral RNA for weeks after the resolution of symptoms.<sup>28</sup>

Issuing binary results may create confusion by equating a sample with a very high viral load to one with a significantly lower viral load.<sup>29</sup> Whilst Ct values from patient samples have demonstrated clinical utility in certain scenarios,<sup>30</sup> including Ct values routinely on a PCR result report could be misleading to the clinician.<sup>31</sup> It is thus critical for the requesting healthcare worker to discuss the PCR result with the clinical virologist whilst considering a patient's history, contacts, symptoms, radiological features and previous laboratory results. From a laboratory perspective, thorough knowledge of the performance characteristics of individual PCR assays in use is necessary to accurately interpret results.<sup>32</sup> Furthermore, a sample tested on different assays can yield different Ct values reflecting differences in the targets detected and the chemistry of the test used. Consequently, when considering trends in Ct values it is preferable to test samples with the same assay or platform each time.

## Conclusion

The COVID-19 pandemic has prompted a rapid upscaling of diagnostic platforms in both private and public laboratories. Whilst the PCR test currently remains the gold standard for the diagnosis of COVID-19, it is not without challenges. Pre-analytical, analytical and postanalytical factors may influence the outcome of a result. Alternative diagnostic algorithms incorporating immunoassays and novel molecular methods are currently under investigation.

## Acknowledgements

### Competing interests

The authors declare that they have no financial or personal relationships that may have inappropriately influenced them in writing this article.

### Authors' contributions

A.B., K.G. and V.B. contributed to writing and editing of the manuscript.

### Ethical considerations

This article followed all ethical standards for research without direct contact with human or animal subjects.

### Funding information

This work is based on the research supported in part by the National Research Foundation of South Africa with a grant (unique grant no. 99238).

Research reported in this publication was supported by the Fogarty International Centre (FIC), National Institutes of Health Common Fund, Office of Strategic Coordination, Office of the Director (NIH CF/OSC/OD), Office of AIDS Research, Office of the Director (OAR/OD/NIH) and the National Institute of Mental Health of the NIH (NIMH/NIH, award no. D43TW010131). The content is solely the responsibility of the authors and does not necessarily represent the official views of the NIH.

## Data availability

Data sharing is not applicable to this article as no new data were created or analysed in this study.

## Disclaimer

The views and opinions expressed in this article are those of the authors and do not necessarily reflect the official policy or position of an affiliated agency of the authors.

## References

- Lippi G, Simundic AM, Plebani M. Potential preanalytical and analytical vulnerabilities in the laboratory diagnosis of coronavirus disease 2019 (COVID-19). *Clin Chem Lab Med*. 2020;58(7):1070–1076. <https://doi.org/10.1515/cclm-2020-0285>
- Wang X, Tan L, Wang X, et al. Comparison of nasopharyngeal and oropharyngeal swabs for SARS-CoV-2 detection in 353 patients received tests with both specimens simultaneously. *Int J Infect Dis*. 2020;94:107–109. <https://doi.org/10.1016/j.ijid.2020.04.023>
- Pan Y, Zhang D, Yang P, Poon LLM, Wang Q. Viral load of SARS-CoV-2 in clinical samples. *Lancet Infect Dis*. 2020;20(4):411–412. [https://doi.org/10.1016/S1473-3099\(20\)30113-4](https://doi.org/10.1016/S1473-3099(20)30113-4)
- Zou L, Ruan F, Huang M, et al. SARS-CoV-2 viral load in upper respiratory specimens of infected patients. *N Engl J Med*. 2020;382(12):1177–1179. <https://doi.org/10.1056/NEJMc2001737>
- Berenger BM, Conly JM, Fonseca K, et al. Saliva collected in universal transport media is an effective, simple and high-volume amenable method to detect SARS-CoV-2. *Clin Microbiol Infect*. 2020;27(4):656–657. <https://doi.org/10.1016/j.cmi.2020.10.035>
- Tsang NNY, So HC, Ng KY, Cowling BJ, Leung GM, Ming DK. Diagnostic performance of different sampling approaches for SARS-CoV-2 RT-PCR testing: A systematic review and meta-analysis. *Lancet Infect Dis*. 2021;21(9):1233–1245. [https://doi.org/10.1016/S1473-3099\(21\)00146-8](https://doi.org/10.1016/S1473-3099(21)00146-8)
- Centres for Disease Control and Prevention. Interim guidelines for collecting and handling of clinical specimens for COVID-19 testing: CDC [homepage on the Internet]. 2021 [cited 2021 Aug 16]. Available from: <https://www.cdc.gov/coronavirus/2019-ncov/lab/guidelines-clinical-specimens.html>
- National Department of Health. Coronavirus disease 2019 (COVID-19) caused by a novel coronavirus (SARS-CoV-2): Guidelines for case-finding, diagnosis, and public health response in South Africa version 3.0. [homepage on the Internet]. 2020 [cited 2022 May 04]. Available from: [https://www.nicd.ac.za/wp-content/uploads/2020/07/NICD\\_DoH-COVID-19-Guidelines\\_Final\\_3-Jul-2020.pdf](https://www.nicd.ac.za/wp-content/uploads/2020/07/NICD_DoH-COVID-19-Guidelines_Final_3-Jul-2020.pdf)
- World Health Organization. Laboratory testing for coronavirus disease 2019 (COVID-19) in suspected human cases: Interim guidance [homepage on the Internet]. Geneva: WHO; 2020 [cited 2021 Aug 16]. Available from: <https://apps.who.int/iris/handle/10665/331329>
- Freire-Paspuel B, Vega-Mariño P, Velez A, et al. Cotton-tipped plastic swabs for SARS-CoV-2 RT-qPCR diagnosis to prevent supply shortages. *Front Cell Infect Microbiol*. 2020;10:356. <https://doi.org/10.3389/fcimb.2020.00356>
- Higgins TS, Wu AW, Ting JY. SARS-CoV-2 nasopharyngeal swab testing-false-negative results from a pervasive anatomical misconception. *JAMA Otolaryngol Head Neck Surg*. 2020;146(11):993–994. <https://doi.org/10.1001/jamaoto.2020.2946>
- Zheng J. SARS-CoV-2: An emerging coronavirus that causes a global threat. *Int J Biol Sci*. 2020;16(10):1678–1685. <https://doi.org/10.7150/ijbs.45053>
- Wang W, Xu Y, Gao R, et al. Detection of SARS-CoV-2 in different types of clinical specimens. *JAMA*. 2020;323(18):1843. <https://doi.org/10.1001/jama.2020.3786>
- Li L, Li X, Guo Z, et al. Influence of storage conditions on SARS-CoV-2 nucleic acid detection in throat swabs. *J Infect*. 2020;222(2):203–205. <https://doi.org/10.1093/infdis/jiaa272>
- Skalina KA, Goldstein DY, Sulail J, et al. Extended storage of SARS-CoV-2 nasopharyngeal swabs does not negatively impact results of molecular-based testing across three clinical platforms. *J Clin Pathol*. 2022;75(1):61–64. <https://doi.org/10.1136/jclinpath-2020-206738>
- Afzal A. Molecular diagnostic technologies for COVID-19: Limitations and challenges. *J Adv Res*. 2020;26:149–159. <https://doi.org/10.1016/j.jare.2020.08.002>
- Sethuraman N, Jeremiah SS, Ryo A. Interpreting diagnostic tests for SARS-CoV-2. *JAMA*. 2020;323(22):2249. <https://doi.org/10.1001/jama.2020.8259>
- Huggett JF, Benes V, Bustin SA, et al. Cautionary note on contamination of reagents used for molecular detection of SARS-CoV-2. *Clin Chem*. 2020;66(11):1369–1372. <https://doi.org/10.1093/clinchem/hvaa214>
- Wernike K, Keller M, Conraths FJ, Mettenleiter TC, Groschup MH, Beer M. Pitfalls in SARS-CoV-2 PCR diagnostics. *Transbound Emerg Dis*. 2021;68(2):253–257. <https://doi.org/10.1111/tbed.13684>
- Gribble J, Stevens LJ, Agostini ML, et al. The coronavirus proofreading exoribonuclease mediates extensive viral recombination. *PLoS Pathog*. 2021;17(1):e1009226. <https://doi.org/10.1371/journal.ppat.1009226>
- Petersen E, Ntoumi F, Hui DS, et al. Emergence of new SARS-CoV-2 variant of concern omicron (B.1.1.529) – Highlights Africa's research capabilities, but exposes major knowledge gaps, inequities of vaccine distribution, inadequacies in global COVID-19 response and control efforts. *Int J Infect Dis*. 2022;114:268–272. <https://doi.org/10.1016/j.ijid.2021.11.040>
- Sheikh A, McMenamin J, Taylor B, Robertson C. SARS-CoV-2 delta VOC in Scotland: Demographics, risk of hospital admission, and vaccine effectiveness. *Lancet*. 2021;397(10293):2461–2462. [https://doi.org/10.1016/S0140-6736\(21\)01358-1](https://doi.org/10.1016/S0140-6736(21)01358-1)
- Zou L, Ruan F, Huang M, et al. SARS-CoV-2 viral load in upper respiratory specimens of infected patients. *N Engl J Med*. 2020;382(12):1177–1179. <https://doi.org/10.1056/NEJMc2001737>
- Cevik M, Tate M, Lloyd O, Maraolo AE, Schafers J, Ho A. SARS-CoV-2, SARS-CoV, and MERS-CoV viral load dynamics, duration of viral shedding, and infectiousness: A systematic review and meta-analysis. *Lancet Microbe*. 2021;1(1):13–22. [https://doi.org/10.1016/S2666-5247\(20\)30172-5](https://doi.org/10.1016/S2666-5247(20)30172-5)
- La Scola B, Le Bideau M, Andreani J, et al. Viral RNA load as determined by cell culture as a management tool for discharge of SARS-CoV-2 patients from infectious disease wards. *Eur J Clin Microbiol Infect Dis*. 2020;39(6):1059–1061. <https://doi.org/10.1007/s10096-020-03913-9>
- Rhee C, Kanjilal S, Baker M, Klompas M. Duration of severe acute respiratory syndrome coronavirus 2 (SARS-CoV-2) infectivity: When is it safe to discontinue isolation? *Clin Infect Dis*. 2020;72(8):1467–1474. <https://doi.org/10.1093/cid/ciaa1249>
- Salvatore PP, Dawson P, Wadhwa A, et al. Epidemiological correlates of polymerase chain reaction cycle threshold values in the detection of severe acute respiratory syndrome coronavirus 2 (SARS-CoV-2). *Clin Infect Dis*. 2020;72(11):761–767. <https://doi.org/10.1093/cid/ciaa1469>
- Xiao AT, Tong YX, Zhang S. Profile of RT-PCR for SARS-CoV-2: A preliminary study from 56 COVID-19 patients. *Clin Infect Dis*. 2020;71(16):2249–2251. <https://doi.org/10.1093/cid/ciaa460>
- Tom MR, Mina MJ. To interpret the SARS-CoV-2 test, consider the cycle threshold value. *Clin Infect Dis*. 2020;71(16):2252–2254. <https://doi.org/10.1093/cid/ciaa619>
- Rao SN, Manissero D, Steele VR, Pareja J. A narrative systematic review of the clinical utility of cycle threshold values in the context of COVID-19. *Infect Dis Ther*. 2020;9:573–586. <https://doi.org/10.1007/s40121-020-00324-3>
- Binnicker MJ. Challenges and controversies to testing for COVID-19. *J Clin Microbiol*. 2020;58(11):e01695-20. <https://doi.org/10.1128/JCM.01695-20>
- Carroll A, McNamara E. Comparison and correlation of commercial SARS-CoV-2 real-time-PCR assays. *Euro Surveill*. 2021;26(6):2002079. <https://doi.org/10.2807/1560-7917.ES.2021.26.6.2002079>